



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Aug 1, 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL

DISTRICT: 8 ADDRESS: SR101 MP77

LOCATION: Artic 77

COUNTY: Grays Harbor

MAKE: CARDINAL

TYPE: Trk Scale

ACCURACY CLASS: IIII

MODEL: 738

ID. NO: 9011-201 D

CAPACITY: 100000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: Digital

PLATFORM 8x10

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| SEC. 1 | 20000 | 20000 | 0 | | | | |
| | | | | | | | |
| CENTER | 20000 | 20000 | 0 | | | | |
| | | | | | | | |
| SEC. 2 | 20000 | 20000 | 0 | | | | |
| | | | | | | | |
| | | | | | | | |
| | | | | | | | |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.


Signed by:

David Cromer

Date:

Aug 20, 2012

Print name: David Cromer



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on August 20, 2012

Patricia A. Geppert
(Signature of notary public)

Printed Name of Notary

My appointment expires Sept 4, 2015